

PART 898 PERFORMANCE STANDARD FOR ELECTRODE LEAD WIRES AND PATIENT CABLES

目录

PART 898 PERFORMANCE STANDARD FOR ELECTRODE LEAD WIRES AND PATIENT CABLES.....	1
Sec. 898.11 Applicability.....	2
Sec. 898.12 Performance standard.....	2
Sec. 898.13 Compliance dates.....	2
Sec. 898.14 Exemptions and variances.....	3

Sec. 898.11 Applicability.

Electrode lead wires and patient cables intended for use with a medical device shall be subject to the performance standard set forth in 898.12.

Sec. 898.12 Performance standard.

(a) Any connector in a cable or electrode lead wire having a conductive connection to a patient shall be constructed in such a manner as to comply with subclause 56.3(c) of the following standard:

International Electrotechnical Commission (IEC)

601-1: Medical Electrical Equipment

601-1 (1988) Part 1: General requirements for safety

Amendment No. 1 (1991)

Amendment No. 2 (1995).

(b) Compliance with the standard shall be determined by inspection and by applying the test requirements and test methods of subclause 56.3(c) of the standard set forth in paragraph (a) of this section.

Sec. 898.13 Compliance dates.

The dates for compliance with the standard set forth in 898.12 (a) shall be as follows:

(a) For electrode lead wires and patient cables used with, or intended for use with, the following devices, the date for which compliance

is required is May 11, 1998:

Listing of Devices for Which Compliance is Required Effective

May 11, 1998

Phase	Product code	21 CFR section	Class	Device name
1	73 BZQ	868.2375	II	Monitor, Breathing Frequency.
1	73 FLS	868.2375	II	Monitor (Apnea Detector), Ventilatory Effort.
1	74 DPS	870.2340	II	Electrocardiograph.
1	74 DRG	870.2910	II	Transmitters and Receivers, Physiological Signal, Radio Frequency.
1	74 DRT	870.2300	II	Monitor, Cardiac (including Cardiotachometer and Rate Alarm).
1	74 DRX	870.2360	II	Electrode, Electrocardiograph.
1	74 DSA	870.2900	II	Cable, Transducer and Electrode, Patient (including Connector).
1	74 DSH	870.2800	II	Recorder, Magnetic Tape, Medical.
1	74 DSI	870.1025	III	Detector and Alarm, Arrhythmia.
1	74 DXH	870.2920	II	Transmitters and Receivers, Electrocardiograph, Telephone.

(b) For electrode lead wires and patient cables used with, or intended for use with, any other device, the date for which compliance is required is May 9, 2000.

Sec. 898.14 Exemptions and variances.

(a) A request for an exemption or variance shall be submitted in the form of a petition under 10.30 of this chapter and shall comply with the requirements set out therein. The petition shall also contain the

following:

(1) The name of the device, the class in which the device has been classified, and representative labeling showing the intended uses (s) of the device;

(2) The reasons why compliance with the performance standard is unnecessary or unfeasible;

(3) A complete description of alternative steps that are available, or that the petitioner has already taken, to ensure that a patient will not be inadvertently connected to hazardous voltages via an unprotected patient cable or electrode lead wire for intended use with the device; and

(4) Other information justifying the exemption or variance.

(b) An exemption or variance is not effective until the agency approves the request under 10.30 (e) (2) (i) of this chapter.



电话: 0755-86664989



QQ: 2539440171



邮箱: consultant@hlongmed.com



龙德医疗器械资讯

丰富的医疗器械课程直播平台
专业的医疗器械服务平台
齐全的医疗器械交流社区



医课汇

专业的医疗器械医课课程平台
全面的医疗器械资讯平台
专业的医疗器械人才培养

专业带去价值, 服务赢来美誉!

邮箱: consultant@hlongmed.com

网址: www.hlongmed.com