

MDCG 2022-10

Q&A on the interface between Regulation (EU) 536/2014 on clinical trials for medicinal products for human use (CTR) and Regulation (EU) 2017/746 on *in vitro* diagnostic medical devices (IVDR)

May 2022

This document has been endorsed by the Medical Device Coordination Group (MDCG) and the Clinical Trial Expert Group (CTEG). The MDCG and the CTEG are composed of representatives of all Member States and are chaired by a representative of the European Commission.

The document is not a European Commission document and it cannot be regarded as reflecting the official position of the European Commission. Any views expressed in this document are not legally binding and only the Court of Justice of the European Union can give binding interpretations of Union law.

Introduction

This Q&A intends to clarify certain interfaces between the Regulation (EU) No 536/2014 on clinical trials for medicinal products for human use (CTR) and Regulation (EU) 2017/746 on *in vitro* diagnostic medical devices (IVDR)¹. It was developed by clinical trials experts from Clinical Trials Facilitation and Coordination Group (CTFG) and *in vitro* diagnostics experts from the IVD sub-group of the Medical Device Coordination Group (MDCG)². It has been adopted by CTFG, the Clinical Trials Expert Group of the European Commission and endorsed by the MDCG.

The need for interoperability of databases for clinical trials with medicinal products and medical devices is laid down in the new Regulations, however, no further details on the interface of IVDR and CTR are in the legislation, e.g. the requirements for assays used in the context of clinical trials.

Assays used in clinical trials may range from CE marked *in vitro* diagnostic medical devices (IVDs) to trial- or medicinal product-specific assays that are not always meant to be developed as IVDs. The need to clarify requirements for these assays lead to the Q&A and the concept of the “medical purpose of an assay in a clinical trial” as a specific setting.

The CTR created a new framework for the assessments of initial clinical trial applications and applications for substantial modifications. The assessment procedure under the CTR will be driven by strict timelines and will require close cooperation between Member States concerned in multi-country trials. The aim of this Q&A is to support these coordinated assessments by providing clarification on the regulatory status of assays performed on human samples used in the context of clinical trials as well as on the regulatory expectations toward the clinical trial sponsors. The overall aim is to support the conduct of clinical trials using diagnostic assays, including combined trials for the development of companion diagnostics (CDx).

Non-interventional studies, defined as clinical studies other than a clinical trial in Article 2 (4) of the CTR are out of scope of both the CTR (as per Article 1) and this document. However, IVDs used in non-interventional clinical studies are subject to the IVDR.

The IVDR lays down rules concerning the “placing on the market”, “making available on the market” or “putting into service” of IVDs and accessories for an IVD. It further outlines the requirements for performance studies concerning IVDs and accessories for an IVD conducted in the Union.

¹ [Regulation \(EU\) 2017/746](#) (OJ L 117, 5.5.2017, p. 176) replaces [Directive 98/79/EC](#) (OJ L 331, 7.12.1998, p. 1) on *in vitro* diagnostic medical devices as of 26 May 2022. Regulation (EU) 2017/746 maintains many of the legislative principles of Directive 98/79/EC e.g. regarding general definitions and CE-marking after conformity assessment before placing medical devices on the EU market. Regulation (EU) 2017/746 introduces several important changes including new definitions, Unique Device Identification (UDI) system for medical devices, and new principles of classification of IVDs. Please note that [Regulation \(EU\) 2022/112](#) (OJ L 19, 28.1.2022, p. 3) amends Regulation (EU) 2017/746 with respect to transitional provisions for certain *in vitro* diagnostic medical devices and the deferred application of conditions for in-house devices.

² Expert group responsible for overseeing the implementation of Regulation (EU) 2017/746, set up according to Art. 103 of Regulation (EU) 2017/745 and Art. 98 of Regulation (EU) 2017/746.

When reference is made to IVDs in the following text, this also includes CDx, as the requirements do not differ from a clinical trials perspective. Information is included on the use of in-house IVDs in the context of clinical trials.

The questions and answers below specifically addresses the use of assays in the framework of clinical trials conducted under the CTR, in line with requirements of the IVDR and should not be understood to apply beyond it.

The same issue might be covered in multiple questions, the reader is therefore encouraged to read the entire document.

For guidance documents issued by the MDCG regarding IVDs refer to [the European Commission website](#).

For guidance on clinical trials please refer to [EudraLex Vol. 10](#).

Glossary

CE marking: According to Article 2 (35) IVDR, 'CE marking of conformity' or 'CE marking' means a marking by which a manufacturer indicates that a device is in conformity with the applicable requirements set out in this Regulation and other applicable Union harmonisation legislation providing for its affixing

Performance Study Plan: According to Article 2 (43) IVDR 'performance study plan' means a document that describes the rationale, objectives, design methodology, monitoring, statistical considerations, organisation and conduct of a performance study;

Clinical Trial Application (CTA): According to CTR Article 5 or Article 11 an application for authorisation submitted to the EU Clinical Trials Portal and Database

Clinical Trial Regulation (CTR): [Regulation \(EU\) No 536/2014](#) (CTR) of 16 April 2014 on clinical trials on medicinal products for human use, and repealing Directive 2001/20/EC, OJ L 158, 27.5.2014, p. 1. The expected application date of the CTR is 30 January 2022.

Companion diagnostic (CDx): Article 2 (7) IVDR' means a device which is essential for the safe and effective use of a corresponding medicinal product to:

- (a) identify, before and/or during treatment, patients who are most likely to benefit from the corresponding medicinal product; or
- (b) identify, before and/or during treatment, patients likely to be at increased risk of serious adverse reactions as a result of treatment with the corresponding medicinal product;

Combined Trial: For the purpose of this document a combined trial is a simultaneous investigation of a medicinal product (clinical trial authorised under the CTR) and an IVD (clinical performance study). The combined trial is subject to requirements of both the Clinical trial legislation (CTR when this applies) and the IVD legislation (IVDR when this applies).

Device for performance study according to IVDR Article 2 (45) means a device intended by the manufacturer to be used in a performance study. A device intended to be used for research purposes, without any medical objective, shall not be deemed to be a device for performance study

Stratification: Within clinical trials, a method used in the randomization to ensure equal distribution of chosen variables between treatment arms

In-house IVD: An IVD manufactured and used within the same health institution as outlined in IVDR Article 5 (5). Health institution is defined in IVDR Article 2 (29).

Instructions for use according to Article 2 (14) IVDR means the information provided by the manufacturer to inform the user of the device's intended purpose and proper use and of any precautions to be taken

Intended purpose according to Article 2 (12) IVDR means the use for which a device is intended according to the data supplied by the manufacturer on the label, in the instructions for use or in promotional or sales materials or statements or as specified by the manufacturer in the performance evaluation

Interventional clinical performance study according to Article 2 (46) IVDR means a clinical performance study where the test results may influence patient management decisions and/or may be used to guide treatment

Clinical Trial according to Article 2 (2) CTR, means a clinical study which fulfils any of the following conditions: (a) the assignment of the subject to a particular therapeutic strategy is decided in advance and does not fall within normal clinical practice of the Member State concerned; (b) the decision to prescribe the investigational medicinal products is taken together with the decision to include the subject in the clinical study; or (c) diagnostic or monitoring procedures in addition to normal clinical practice are applied to the subjects. By definition, all clinical trials are “interventional” For clarification on the differentiation between clinical trial and non-interventional study, reference is also made to Question 1.5 in the [Q&A published on EudraLex vol. 10](#).

In vitro diagnostic medical device (IVD) is a medical device as defined in point (1) of Article 2 of Regulation (EU) 2017/745 on medical devices and, more specifically, according to Article 2 (2) IVDR, any medical device which is a reagent, reagent product, calibrator, control material, kit, instrument, apparatus, piece of equipment, software or system, whether used alone or in combination, intended by the manufacturer to be used in vitro for the examination of specimens, including blood and tissue donations, derived from the human body, solely or principally for the purpose of providing information on one or more of the following: (a) concerning a physiological or pathological process or state; (b) concerning congenital physical or mental impairments; (c) concerning the predisposition to a medical condition or a disease; (d) to determine the safety and compatibility with potential recipients; (e) to predict treatment response or reactions; (f) to define or monitoring therapeutic measures. Specimen receptacles shall also be deemed to be in vitro diagnostic medical devices;

IVD manufacturer according to Article 2 (23) IVDR means a natural or legal person who manufactures or fully refurbishes a device or has a device designed, manufactured or fully refurbished, and markets that device under its name or trade mark

IVDR: [Regulation \(EU\) 2017/746](#) of the European Parliament and of the Council of 5 April 2017 on *in vitro* diagnostic medical devices and repealing Directive 98/79/EC and Commission Decision 2010/227/EU, OJ L 117, 5.5.2017, p. 176.

Making available on the market according to Article 2 (20) IVDR, means any supply of a device, other than a device for performance study, for distribution, consumption or use on the Union market in the course of a commercial activity, whether in return for payment or free of charge;

Performance study according to Article 2 (42) IVDR, means a study undertaken to establish or confirm the analytical or clinical performance of a device

Placing on market according to Article 2 (21) IVDR, means the first making available of a device, other than a device for performance study, on the Union market

Putting into service according to Article 2(22) IVDR, means the stage at which a device, other than a device for performance study, has been made available to the final user as being ready for use on the Union market for the first time for its intended purpose;

Substantial modification according to Article 2.2.13 CTR means any change to any aspect of the clinical trial which is made to ongoing clinical trials and which is likely to have a substantial impact on the safety or rights of the subjects or on the reliability and robustness of the data generated in the clinical trial.

Questions and Answers

1. **Question: Regarding IVDs that are currently on the market on the basis of Directive 98/79/EC, do they need to undergo conformity assessment under the IVDR to remain on the market under that Regulation?**

Answer: The IVDR does not foresee a recognition of Directive marketed IVDs but lays down various transitional provisions for IVDs compliant with Directive 98/79/EC. For more details on transitional provisions see the [EU Commission website](#).

2. **Question: Do all IVDs used in a clinical trial require CE marking at the time of the clinical trial?**

Answer: Not all. An IVD used in a clinical trial needs to be compliant with the *in vitro* diagnostic medical device legislation. This implies that the IVD either has the CE marking for the intended purpose, or is an in-house IVD (see Q3) or is a device for a performance study according to the IVDR ongoing in parallel (see Q6, Q7). In-house IVDs and devices for performance study are not required to bear CE-marking.

3. Question: Can I use an in-house IVD in a clinical trial?

Answer: Yes. This remains possible.

With the exception of the relevant general safety and performance requirements set out in Annex I, the requirements of the IVDR shall not apply to devices manufactured and used only within health institutions established in the Union, provided that all of the conditions of Article 5(5) are met.

In-house IVDs might also be regulated in national legislation and it is necessary to check with the national competent authorities responsible for performance studies, which requirements apply.

4. Question: Are all assays used in a clinical trial subject to the IVD legislation?

Answer: No, not all assays, but those that fulfil the definition of an IVD are subject to the IVD legislation. Such assays may have a medical purpose within the clinical trial, e.g. when they guide medical management decisions or follow-up (see Q6).

The IVD Directive 98/79/EC and the IVDR lay down rules concerning the placing on the market, making available on the market or putting into service of IVD for human use and accessories for such devices in the European Union. In addition, the IVDR lays down rules specific to devices for performance studies and in-house devices.

5. Question: When does an assay used on human samples qualify as an IVD in a clinical trial?

Answer: An assay is considered an IVD if the manufacturer assigns an intended purpose that fulfils the definition of an IVD according to IVDR Article 2 (see Q6). Where a clinical trial sponsor assigns a medical purpose to an assay in the context of the clinical trial in a way that the assay fulfils the definition of an IVD according to IVD Regulation 2017/746 Article 2, the clinical trial sponsor may assume the role of a manufacturer under the IVDR³. In this role, it is up to the clinical trial sponsor to determine the regulatory status of the assay based on the planned use in the clinical trial. In case of doubt, consultation of the respective NCAs for IVDs and for clinical trials (with medicinal products) is recommended.

³ See Article 16(1) IVDR.

6. Question: What are examples of assays used in clinical trials which are IVDs?

Answer: Some examples include assays providing information for clinical trial related medical management decisions (typically to select patients for enrolment in the trial, assign patients to a treatment arm, etc.) and/or may be used to guide follow up measures during and beyond the clinical trial. This would, for example, not be the case, in settings where all trial participants are tested irrespective of treatment arm or medical management and the analysis of impact is conducted retrospectively and where medical management is not impacted by assay results.

Figure 1 visualises, as an example, the flow of a blinded clinical trial with two treatment arms, where the key processes for which assays might be utilised are highlighted. The processes in blue are considered to be used for medical management decisions of trial subjects. These include assays used for inclusion and exclusion of subjects, treatment allocation as well as monitoring the safety and efficacy of the treatment during the trial.

The processes in pink are likely not to impact the medical management of the trial subjects. These include stratification and endpoint analysis or other exploratory assays for which correlation with clinical parameters is investigated retrospectively without impact on patient treatment (medical purpose). In relation to endpoints, it is important to acknowledge that these assays may be considered IVDs in future clinical trials (e.g. used for allocation or monitoring). Where this development is predictable, the assay should be developed and validated in compliance with the applicable requirements of Annex I of the IVDR as an IVD from the beginning. Importantly, in most cases, the assay will also be utilised during the trial as part of the monitoring of the trial subjects, which implies need for compliance with IVDR requirements.




Fig. 1 Simplified examples of use of assays on human samples in a clinical trial. Assays marked in blue (diamonds) are considered to be assays which will likely be considered IVDs as they are used for medical management decisions of trial subjects within the trial. The processes in yellow pink (ellipses) are considered to likely not to impact the medical management of the trial subjects and therefore would not have a medical purpose in the trial.

7. Question: In which situations may a non-CE marked IVD (or an IVD that is CE marked for another intended purpose) be used in a clinical trial for a medical purpose according to the IVDR?

Answer: Note that using a CE marked IVD outside its prescribed intended purpose is not covered by the CE marking provided to that device and renders it a non-CE marked IVD in the context of this clinical trial.

Where a non-CE marked IVD is used for a medical purpose in a clinical trial the IVD must either be

- a device for performance study (it is not acceptable to use an investigational IVD without evaluating its performance), or
- an in-house IVD (see also Q3).

If a non-CE marked assay is used for a medical purpose in a way that the assay fulfils the definition of an IVD according to Article 2 of the IVDR, outside of the abovementioned settings, this is not compliant with the IVDR. It is the responsibility of the clinical trial sponsor to ensure compliance with relevant provisions of Union and national law, including the IVDR.

If performance data is generated for an IVD outside of a performance study it might be insufficient for a future performance evaluation, and thus further studies might be required to generate sufficient clinical evidence.

8. Question: Is a use in research sufficient justification for an assay to be defined as Research Use Only?

Answer: No. Where an assay, instrument, apparatus, appliance material or other article, including software is intended for research use only and is assigned a medical purpose by the clinical trial sponsor in the clinical trial protocol in a way that the assay fulfils the definition of an IVD according to IVDR Article 2, it becomes an IVD and can no longer be considered a research use only assay. It becomes regulated under the IVDR (see Q4-Q6).

9. Question: Can an EU core lab facility⁴ that is part of a health institution established in the Union and has developed an in-house IVD receive samples from several clinical trial sites for analysis, or does the in-house IVD device have to be co-located with the clinical trial site?

Answer: In-house devices can be used as long as the provisions of the IVDR and relevant national legislations are satisfied, which include notably, that the device must be manufactured and used within a health institution established in the EU and the device is not transferred to another legal entity. In-house

⁴ A central laboratory to analyse samples from multiple sites within a clinical trial

IVD must fulfil relevant general safety and performance requirements set out in Annex I.⁵

Samples can travel and be analysed by means of in-house devices based in a different location than the trial site facilitating a core lab facility approach. The in-house concept applies to the device itself, which must be manufactured and used within the responsible health institution.

10. Question: Is there a common EU procedure for combined trials?

Answer: No, there is no dedicated procedure foreseen in the Regulations and therefore, currently, no harmonized procedure is in place. However, where a single combined trial will serve as the clinical performance study for the device and the clinical trial of the medicinal product, the study documentation required for submissions (if applicable) may be partially overlapping. Not all clinical performance studies have to be submitted to the regulatory authorities (see Q14). Sponsors of these clinical trials are encouraged to consult national guidance documents and contact National Competent Authorities prior to clinical trial submission.

11. Question: Which IVDs should be listed in the cover letter of the clinical trial application?

Answer: Only those IVDs should be mentioned in the cover letter that are specifically required by the protocol to be used to achieve the objectives of the clinical trial. The information provided should permit identification of the IVD being used and whether it is CE-marked for the planned use. Evidence of CE-marking generally does not have to be provided in the CTA, but it is the responsibility of the clinical trial sponsor to determine whether the device is used in the clinical trial in line with the manufacturer's intended purpose.

12. Question: Which information needs to be provided in the clinical trial application⁶, if an assay with medical purpose that qualifies as an IVD is used in a clinical trial where the assay does not have a CE marking for the intended purpose (including in-house assays)?

Answer: According to Annex I.B.7 (i) of the CTR, the clinical trial sponsor shall include in the Cover letter the list of medical devices which are to be investigated in the clinical trial but which are not part of the investigational medicinal product or products, together with a statement as to whether the

⁵ The health institution must provide reasoned justification on which requirements are not fully met according to article 5(f) (iii) IVDR.

⁶ In this document 'clinical trial application' covers both the initial trial application and applications for substantial modifications to add an assay with a medical purpose to an ongoing trial

medical devices are CE-marked for the intended purpose. This should be understood to refer to medical devices (including IVDs) which are specifically required by the protocol to be used to achieve the objectives of the trial.

Use of an IVD in a clinical trial may meet the definition of a device for performance study. (See glossary 'device for performance study' and 'performance study')

For each device used in the clinical trial that meets the definition of a device for performance study, the sponsor of the clinical trial should comment on the need for notification or application for the clinical performance study conducted in parallel with the clinical trial, based on the applicable IVD legislation. Reference should be made to EUDAMED Single Identification Number (SIN) and/or National Competent Authority reference numbers (if available) for planned or submitted performance studies. To facilitate bridging between clinical trial and performance study, the IVD being used should be identified in the cover letter and protocol of the clinical trial.

Where the IVD manufacturer is supporting the clinical trial, which is also a performance study, the clinical trial sponsor should obtain a statement from the IVD manufacturer that the device for performance study in question conforms to the general safety and performance requirements laid down in Annex I of the IVDR, apart from the aspects covered by the clinical performance study and that, with regards to those aspects, every precaution has been taken to protect the health and safety of the subject.

Where the clinical trial sponsor is also the manufacturer of the IVD or assumes the role as manufacturer of the IVD according to Article 16 IVDR, the clinical trial sponsor must draw up their own statement as above. In case the study falls under IVDR Art 58 (1) or (2), it must be designed, authorised, conducted, recorded and reported in accordance with IVDR Art. 58-77 and Annex XIV.

In the CTA cover letter, the availability of this statement should be confirmed, and the aspects that are covered in performance study as part of the clinical trial should be listed. The statement is to be kept with the sponsor's trial master file (TMF) and investigators' site file (ISF) and made available for inspections or at the request of NCAs.

In case of confidential information in the description of the performance studies, a redacted cover letter can be submitted, as necessary.

In addition, further details on compliance with IVDR and Annex I of the IVDR can be provided in the protocol, according to CTR Annex I 17. a)-f).

13. Question: Which information needs to be provided in the clinical trial application for an in-house IVD?

Answer: In principle, in the case of in-house IVDs, the sponsor should provide the name of the health institution where the IVD is manufactured and used and a link to their IVDR Art 5(f) declaration. It is recommended to document this aspect in the harmonised template document⁷.

The IVD analysis must be carried out in a laboratory complying with ISO15189, or, where applicable, with national provisions, in line with IVDR Article 5(5)(c).

In addition, further details on compliance with IVDR Article 5(5) and Annex I of the IVDR can be provided in the protocol, according to CTR Annex I 17. a)-f).

Please note that [Regulation \(EU\) 2022/112](#) amends the IVDR to defer the application of certain conditions for in-house devices.

14. Question: What are the notification/application requirements for the performance study according to the IVDR?

Answer: Where the performance study is either an interventional clinical performance study or involves risks for the subjects of the study an application is necessary. Additional reporting and documentation obligations apply according to Art. 58 (1) and (2) IVDR.

Where the IVD is being used according to its intended purpose outlined in the instructions for use, Article 70 may apply. The submission requirements from the IVDR perspective will depend on the particulars of the device in question and the design of the performance study (Article 57, 58 and 70 IVDR).

15. Question: What are the respective responsibilities of the clinical trial sponsor and IVD manufacturer including in the case of co-development and use in a combined trial?

Answer: Clinical trial sponsor requirements are defined in Article 2.2.14 CTR: The sponsor is responsible for the initiation and management of the clinical trial, including the selection and use of IVDs. The sponsor is responsible for the overall compliance of products used in the clinical trial with the CTR and other relevant EU and National legislation. Clinical trial sponsors are reminded of their obligations according to ICH-GCP E6 (R2) to document the competence of a laboratory to perform a certain test to support the reliability of results. This information should be kept at the sponsor's Trial Master File (TMF) and, when applicable, at the investigator's site file. Site suitability in

⁷ Available on [EudraLex vol. 10](#).

accordance with Art 50 and Annex I N67 of the CTR should be documented. As good practice recommendation, this should take place by using the harmonised template document.

In case of a clinical trial where the IVD is used outside its intended purpose the clinical trial sponsor assumes the obligations incumbent on manufacturers according to Article 16(1) of the IVDR⁸.

It is the intended purpose, which determines the applicability of the IVDR for an assay and not the intention (or not) of obtaining a CE-mark at a later stage. Where the assay qualifies as an IVD under the IVDR a natural or legal person⁹ needs to take responsibility for the safety and performance of the IVD device.

16. Question: In situations where it is procedurally feasible and where the same ethics committee is responsible for evaluating the clinical trial AND performance study aspects of a combined trial – Would there be a legal requirement for two separate ethics opinions, or could the outcome be a single document for both aspects?

Answer: According to IVDR Art. 2 (59) and Art 58 (3), and CTR Art. 2.2.11, recital (18) and Art. 4, the ethics committees operate according to national law. It is up to the Member States to organize the work of the ethics committees and it is up to the Member States to ensure that the procedures for the review by the ethics committees are compatible with the clinical trial or conformity assessment procedures in the CTR and IVDR respectively. However, in situations where it is procedurally feasible and where the same ethics committee is responsible for evaluating the clinical trial and performance evaluation aspects of a combined trial, there is no legal requirement for two separate ethics opinions. The outcome could be a single document for both aspects, provided this is permissible under national law of the member state.

At the same time, in accordance with Art 9 of the CTR, Member States shall ensure that the persons validating and assessing the application do not have conflicts of interest, are independent of the sponsor, of the clinical trial site and the investigators involved and of persons financing the clinical trial, as well as free of any other undue influence. Member States shall ensure that persons admitting and assessing the application have no financial or personal interests, which could affect their impartiality.

⁸ See also Q5.

⁹ Which might be the product, a health institution,



医课汇
公众号
专业医疗器械资讯平台
WECHAT OF
HLONGMED



hlongmed.com
医疗器械咨询服务
MEDICAL DEVICE
CONSULTING
SERVICES



医课培训平台
医疗器械任职培训
WEB TRAINING
CENTER



医械宝
医疗器械知识平台
KNOWLEDG
ECENTEROF
MEDICAL DEVICE



MDCPP.COM
医械云专业平台
KNOWLEDG
ECENTEROF MEDICAL
DEVICE