
Non-invasive sphygmomanometers —
Part 2:
Clinical investigation of intermittent
automated measurement type
AMENDMENT 1

Sphygmomanomètres non invasifs —

*Partie 2: Investigation clinique pour type ponctuel à mesurage
automatique*

AMENDEMENT 1



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO and IEC shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents) or the IEC list of patent declarations received (see <http://patents.iec.ch>).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared jointly by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care*, and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC D, *Electromedical equipment*.

A list of all parts in the ISO 81060 series and in the IEC 81060 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Non-invasive sphygmomanometers —

Part 2:

Clinical investigation of intermittent automated measurement type

AMENDMENT 1

Clause 3

Add the following after 3.4:

3.5

TOTAL LIMB CIRCUMFERENCE RANGE

range, from the smallest limb circumference to the largest limb circumference, intended by the MANUFACTURER for use with the AUTOMATED SPHYGMOMANOMETER

5.1.4

Replace the text with the following, and renumber the subsequent formulae:

5.1.4 * Limb size distribution

- a) Limb circumferences shall be distributed as follows:
- 1) at least 20 % of the subjects shall have a limb circumference which lies within each quarter of the TOTAL LIMB CIRCUMFERENCE RANGE;
 - 2) at least 10 % of the subjects shall have a limb circumference which lies within the highest octile of the TOTAL LIMB CIRCUMFERENCE RANGE; and
 - 3) at least 10 % of the subjects shall have a limb circumference within the lowest octile of the TOTAL LIMB CIRCUMFERENCE RANGE.
- b) Additionally, for a SPHYGMOMANOMETER intended for use with multiple CUFF sizes, each CUFF shall be tested on at least N_{cuff} subjects as calculated according to [Formula \(1\)](#).

$$N_{\text{cuff}} = \frac{r_{\text{cuff}}}{2 \cdot r_{\text{total}}} \cdot N_{\text{total}} \quad (1)$$

where

N_{total} is the total number of subjects in the study;

r_{cuff} is the size of the limb circumference range for the individual cuff;

r_{total} is the size of the TOTAL LIMB CIRCUMFERENCE RANGE.

ISO 81060-2:2018/Amd.1:2020(E)

- c) The CLINICAL INVESTIGATION REPORT shall provide the specified arm circumference range of each CUFF in centimetres (inches).
- d) The CLINICAL INVESTIGATION REPORT shall include plots showing all subject results by arm circumference, where the Y axis shows the difference between SPHYGMOMANOMETER-UNDER-TEST and REFERENCE BLOOD PRESSURE VALUES and the X axis shows the actual arm circumference of the subjects.
- e) These plots should also indicate the borders of each CUFF with vertical lines.
- f) Plots shall be provided for both:
 - 1) SYSTOLIC BLOOD PRESSURE; and
 - 2) DIASTOLIC BLOOD PRESSURE.

Check compliance by inspection of the CLINICAL INVESTIGATION REPORT.

5.2.3, list item h)

Replace the text with the following:

- h) *CUFFS for the reference SPHYGMOMANOMETER shall have:
 - 1) a bladder length of 75 % to 100 % of the upper arm circumference;
 - 2) a bladder width of 37 % to 50 % of the upper arm circumference; and
 - 3) a 2-piece construction comprising:
 - i) a distensible inner BLADDER; and
 - ii) a non-distensible outer sleeve.

Check compliance by inspection of the CLINICAL INVESTIGATION REPORT.

6.1.4

Replace the text with the following:

6.1.4 * Limb size distribution

- a) Limb circumferences shall be distributed as follows:
 - 1) at least 20 % of the subjects shall have a limb circumference which lies within each quarter of the TOTAL LIMB CIRCUMFERENCE range;
 - 2) at least 10 % of the subjects shall have a limb circumference which lies within the highest octile of the TOTAL LIMB CIRCUMFERENCE range; and
 - 3) at least 10 % of the subjects shall have a limb circumference within the lowest octile of the TOTAL LIMB CIRCUMFERENCE range.
- b) Additionally, for a SPHYGMOMANOMETER intended for use with multiple CUFF sizes, each CUFF shall be tested on at least N_{cuff} subjects as calculated according to [Formula \(1\)](#).
- c) The CLINICAL INVESTIGATION REPORT shall provide the specified arm circumference range of each cuff in centimetres (inches).

- d) The CLINICAL INVESTIGATION REPORT shall include plots showing all subject results by arm circumference, where the Y axis shows the difference between SPHYGMOMANOMETER-UNDER-TEST and REFERENCE BLOOD PRESSURE values and the X axis shows the actual arm circumference of the subjects.
- e) These plots should also indicate the borders of each cuff with vertical lines.
- f) Plots shall be provided for both:
 - 1) SYSTOLIC BLOOD PRESSURE; and
 - 2) DIASTOLIC BLOOD PRESSURE.

Check compliance by inspection of the CLINICAL INVESTIGATION REPORT.

Annex C

Add the following entries:

bladder	ISO 81060-1:2007, 3.2
total limb circumference range	3.5

